

Integrated Optics, UAB Company code: 302833442 VAT No: LT100007179012 https://integratedoptics.com info@integratedoptics.com



PART NUMBER 0785L-34A ITEM NAME 785 NM LASER (HP DIODE; MM FIBER)

PRODUCT DATASHEET



DESCRIPTION

High power 785 nm laser module with MM fiber.


By default, this type of laser is built with FC/PC connector, but other fiber terminations are available upon request. Details about non-standard connector and the fiber used with it should be discussed with the Integrated Optics sales team.

SPECIFICATIONS

Specifications updated: 12 May 2021


Parameter	Minimum Value	Typical Value	Maximum Value
Central Wavelength, nm	780	785	790
Longitudinal modes	-	Multiple	-
Spectral line width FWHM, nm	-	1	2
Output power, mW	-	1500 ¹	-
Power stability, % (RMS, 8 hrs)	-	1 ²	2
Power stability, % (peak-to-peak, 8 hrs)	-	2 ³	3
Intensity noise, % (RMS, 20 Hz to 20 MHz)	-	0.25 4	0.6
Transversal modes	-	Multiple	-
Control interface type	-	UART ⁵	-
Operation mode	-	APC (CW)	-
Modulation bandwidth, kHz	-	1 6	-
Input voltage, VDC	4.8	5	5.3
External power supply requirement	-	+5 V DC, 1.5 A	-
Dimensions, mm	-	50 x 30 x 18 ⁷	-
Fiber Length, m	0.95	1	1.1
Heat-sinking requirement, °C/W	-	1	-
Optimum heatsink temperature, °C	15	20	30
Warm up time, mins (cold start)	0.1	0.5	1
Temperature stabilization	-	Internal TEC	-
Overheat protection	-	Yes	-
Storage temperature, °C (non-condensing)	-10	-	50
Net weight, kg	0.1	0.12	0.14

TYPICAL SPECTRUM




Typical spectrum of 0785 nm diode laser. Measured with 20 pm resolution.

TYPICAL NEAR FIELD




Max. power consumption, W	5	10 20
Warranty, months (op. hrs)	-	14 (10000) ⁸ -
RoHS	-	Yes -
CE compliance	-	- General - Product Safety Directive (GPSD) 2001/95/EC - (EMC) Directive 2004/108/EC
Laser Safety Class	-	4 -
OEM lasers are not compliant with	-	IEC60825 1:2014 (compliant using additional accessories)
Country of origin	-	Lithuania -



²The long term power test is carried out at constant laser body temperature (+/-0.1 °C) using an optical power meter with an input bandwidth of 10 Hz. The actual measurement rate has a period of about 20 seconds to 1 minute.

Note: Product specifications are subject to change without prior notice to improve reliability, function or design or otherwise.

DRAWING



³ The long term power test is carried out at constant laser body temperature (+/-0.1 °C) using an optical power meter with an input bandwidth of 10 Hz. The actual measurement rate has a period of about 20 seconds to 1 minute.

⁴Noise level is measured with a fast photodiode connected to an oscilloscope. The overall system bandwidth is from 2 kHz to 20 MHz

 $^{^{5}}$ Break-out-boxes AM-C8 and AM-C3 can be used for conversion of UART communication to either USB or RS232.

 $^{^6\,}TTL$ digital modulation up to 10 MHz.

 $^{^{7}\,\}mathrm{Excluding}$ control interface pins and an output window/fiber assembly.

 $^{^{\}rm 8}\,\rm Whichever$ occurs first. The laser has an integrated operational hours counter.